

Infinity® Acute Care System™ - Dräger Evita Infinity® V500

Progressive and comprehensive ventilation therapy



TECHNICAL DATA

Patient type	Adults, pediatrics, neonates
Ventilation settings	
Ventilation mode	VC-CMV VC-SIMV VC-AC VC-MMV PC-CMV PC-BIPAP ¹⁾ / SIMV+ PC-SIMV PC-AC PC-APRV PC-PSV SPN-CPAP/PS SPN-CPAP/VS SPN-CPAP SPN-PPS*
Enhancements	<ul style="list-style-type: none"> - AutoFlow™ / Volume Guarantee – Automatic adaptation of inspiratory flow in volume controlled modes (VC-AC) - Variable Pressure Support - Smart Pulmonary View - ATC™ – Automatic Tube Compensation™ - NIV – Mask Ventilation - SmartCare®/PS 2.0 – Automated clinical protocol in SPN-CPAP/PS - Low Flow PV Loop - O₂-Therapy
Ventilation frequency (RR)	0.5 to 98/min (adults) 0.5 to 150/min (pediatric patients, neonates)
Inspiration time (Ti)	0.11 to 10 s (adults) 0.1 to 10 s (pediatric patients, neonates)
Tidal volume (VT)	0.1 to 3.0 L (adults) under BTPS 0.02 to 0.3 L (pediatric patients) under BTPS 0.002 to 0.1 L (neonates) under BTPS
Inspiratory flow (Flow)	2 to 120 L/min (adults) 2 to 30 L/min (pediatric patients, neonates)



Dräger Evita Infinity® V500



product
design
award

2010

Inspiratory pressure (P _{insp})	1 to 95 mbar (or hPa or cmH ₂ O)
Inspiratory pressure limit (P _{max})	2 to 100 mbar (or hPa or cmH ₂ O)
PEEP / intermittent PEEP (Δ _{int} PEEP)	0 to 50 mbar (or hPa or cmH ₂ O)
Pressure assist (P _{supp})	0 to 95 mbar (or hPa or cmH ₂ O)
Rise time for pressure assist (Slope)	0 to 2 s
O ₂ concentration (FiO ₂)	21 to 100 Vol.%
Trigger sensitivity (Flow trigger)	0.2 to 15 L/min
PC-APRV (optional)	<ul style="list-style-type: none"> - Inspiratory time (Thigh) 0.1 to 30 s - Expiratory time (Tlow) 0.1 to 30 s - Inspiratory pressure (Phigh) 1 to 95 mbar (or hPa or cmH₂O) - Expiratory pressure (Plow) 0 to 50 mbar (or hPa or cmH₂O)
Termination Criteria (expiratory flow) (Exp. term.)	1 to 80 %
Proportional Pressure Support PPS* (optional)	<p>Flow Assist</p> <ul style="list-style-type: none"> - Adults 0 to 30 mbar/L/s (or hPa/L/s or cmH₂O/L/s) - Pediatric patients 0 to 100 mbar/L/s (or hPa/L/s or cmH₂O/L/s) - Neonates 0 to 300 mbar/L/s (or hPa/L/s or cmH₂O/L/s) <p>Volume Assist</p> <ul style="list-style-type: none"> - Adults 0 to 100 mbar/L (or mL/hPa or cmH₂O/L) corresponds to compliance compensation: 10000 to 10 mL/mbar (or mL/hPa or mL/cmH₂O) - Pediatric patients 0 to 1000 mbar/L (or mL/hPa or cmH₂O/L) corresponds to compliance compensation: 1000 to 1 mL/mbar (or mL/hPa or mL/cmH₂O) - Neonates 0 to 4000 mbar/L (or hPa/L or cmH₂O/L) corresponds to compliance compensation: 10000 to 0.5 mL/mbar (or mL/hPa or mL/cmH₂O)
Automatic Tube Compensation ATC™	<p>Inside tube diameter tube Ø</p> <ul style="list-style-type: none"> - Endotracheal tube ET Adults 5 to 12 mm (0.2 to 0.47 inch) Pediatric patients 2 to 8 mm (0.08 to 0.31 inch) - Neonates 2 to 5 mm (0.08 to 0.2 inch) - Tracheostoma tube trach. Adults 5 to 12 mm (0.2 to 0.47 inch) Pediatric patients 2.5 to 8 mm (0.1 to 0.31 inch) - Neonates 2.5 to 5 mm (0.1 to 0.2 inch) <p>Degree of compensation 0 to 100 %</p>
O ₂ Therapy	<p>Continuous Flow 2 to 50 L/min</p> <p>O₂ concentration FiO₂ 21 to 100 Vol%</p>

Measured values displayed

Airway pressure measurement	<p>Plateau pressure P_{plat}</p> <p>Pos. end-exp. pressure PEEP</p> <p>Peak inspiratory pressure PIP</p> <p>Mean airway pressure P_{mean}</p> <p>Min. airway pressure P_{min}</p> <p>Range -60 to 120 mbar (or hPa or cmH₂O)</p>
Flow Measurement	
Minute volume measurement	<p>Total minute volume MV</p> <p>Mandatory minute volume MV_{mand}</p> <p>Spontaneous minute volume MV_{spon}</p> <p>Range 0 to 99 L/min BTPS</p>
Tidal volume measurement	<p>Tidal volume VT</p> <p>Range 0 to 5500 mL BTPS</p>
Frequency measurement	<p>Breathing frequency RR</p> <p>Spontaneous breathing frequency RR_{spon}</p> <p>Range 0/min to 300/min</p>
O ₂ measurement (inspiratory side)	<p>Inspiratory O₂ concentration FiO₂</p> <p>Range 18 to 100 Vol%</p>
CO ₂ measurement in main flow (adults and pediatric patients only)	<p>End-expiratory CO₂ concentration etCO₂</p> <p>Range 0 to 100 mmHg</p>

Computed value displays

Compliance C	Range 0 to 650 mL/mbar (or mL/cmH ₂ O)
Resistance R	Range 0 to 1000 mbar / (L/s) (or cmH ₂ O / (L/s))
Leakage minute volume MVleak	Range 0 to 99 L/min
Rapid Shallow Breathing (RSB)	Range 0 to 9999 (/min/L)
Negative Inspiratory Force (NIF)	Range -80 mbar to 0 mbar (or hPa or cmH ₂ O)
Occlusion pressure P0.1	Range -60 to 130 mbar (or hPa or cmH ₂ O)
Curve displays	Airway pressure Paw (t) -30 to 100 mbar (or hPa or cmH ₂ O) Flow (t) -180 to 180 L/min Volume V (t) 2 to 3000 mL Exp. CO ₂ concentration etCO ₂ 0 to 100 mmHg

Alarms / Monitoring

Expiratory minute volume	High / Low
Airway pressure	High / Low
Insp. O ₂ concentration	High / Low
End-expiratory CO ₂ concentration	High / Low
Tachypnoea monitoring	High
Volume monitoring	High / Low
Apnea alarm time	5 to 60 seconds

Performance data

Control principle	time-cycled, volume-constant, pressure-controlled
Intermittent PEEP duration	1 to 20 expiratory cycles
Medicament nebulisation	for 5, 10, 15, 30 minutes
Bronchial suction	
Disconnection detection	automatic
Reconnection detection	automatic
Oxygen enrichment	max. 3 minutes, adults 100 vol.%, for pediatric patients and neonates factor 1 to 2 from actual FiO ₂ concentration
Active suction phase	max. 2 minutes
Final oxygen enrichment	max. 2 minutes, adults 100 vol.%, for pediatric patients and neonates factor 1 to 2 from actual FiO ₂ concentration
Supply system for spontaneous breathing and P _{supp}	adaptive CPAP system with high initial flow
Inspiratory flow	max 180 L/min
Dead space volume	
with CO ₂ cuvette	<15 mL
without CO ₂ cuvette	<11 mL

Operating data

Mains supply	
Mains power connection	100 V to 240 V, 50/60 Hz
Current consumption	at 230 V max. 1.1 A Ventilation Unit with Medical Cockpit at 230 V max. 1.6 A with GS500 at 100 V max. 2.5 A Ventilation Unit with Medical Cockpit at 100 V max. 3.7 A with GS500
Power consumption	max. 2.5 A Ventilation Unit with Medical Cockpit max. 3.7 A with GS500
in operation, without loading of internal battery	approx. 100 W Ventilation Unit with Medical Cockpit approx. 180 W with GS500
Gas supply	
O ₂ gauge pressure	2.7 to 6.0 bar (or 270 to 600 kPa or 39 to 87 psi)
Air gauge pressure	2.7 to 6.0 bar (or 270 to 600 kPa or 39 to 87 psi)

Physical Specifications

Dimensions (W x H x D)	
Evita Infinity® V500	360 mm x 347 mm x 424 mm (14.17 inch x 13.66 inch x 16.69 inch)
Infinity® C500	414 mm x 284 mm x 95 mm (16.29 inch x 11.18 inch x 3.74 inch)
Evita Infinity® V500 and Infinity® C500	420 mm x 685 mm x 410 mm (16.5 inch x 27.0 inch x 16.1 inch)
Evita Infinity® V500 and Infinity® C500 on trolley	577 mm x 1400 mm x 677 mm (22.7 inch x 55.1 inch x 26.7 inch)
GS500 (mounting on trolley only)	291 mm x 218 mm x 381 mm (11.46 inch x 8.58 inch x 15 inch)
Weight	
Evita Infinity® V500	approx. 16 kg (35.27 lbs)
Infinity® C500	approx. 7 kg (15.43 lbs)
GS500	approx. 10 kg (22 lbs)
Evita Infinity® V500 and Infinity® C500	25 kg (55.1 lbs)
Evita Infinity® V500 and Infinity® C500 on trolley	59 kg (130 lbs)
Mounting: Supporting frame	1,65 kg (3.64 lbs)
Adapter for 38 mm pole	2,35 kg (5.18 lbs)
Diagonal screen size C500	17" TFT color touch screen
Input / Output ports (at Infinity® C500)	<ul style="list-style-type: none"> - 3 external RS232 (9-pin) connectors - 4 USB ports (on the back panel) - 2 USB ports (one on each side panel) - 1 DVI for an independently configurable display with touch screen control (for future use) - 2 DVI (not enabled) - 2 RJ 45 Ethernet connectors

¹⁾ BIPAP, trademark used under license. ATC™, trademarked by Dräger. AutoFlow™, trademarked by Dräger.



HEADQUARTERS

Drägerwerk AG & Co. KGaA
Moislinger Allee 53–55
23558 Lübeck, Germany

www.draeger.com

REGION EUROPE CENTRAL AND EUROPE NORTH

Dräger Medical GmbH
Moislinger Allee 53–55
23558 Lübeck, Germany
Tel +49 451 882 0
Fax +49 451 882 2080
info@draeger.com

REGION EUROPE SOUTH

Dräger Médical S.A.S.
Parc de Haute
Technologie d'Antony 2
25, rue Georges Besse
92182 Antony Cedex, France
Tel +33 1 46 11 56 00
Fax +33 1 40 96 97 20
dlmfr-contact@draeger.com

REGION MIDDLE EAST, AFRICA, CENTRAL AND SOUTH AMERICA

Dräger Medical GmbH
Branch Office Dubai
Dubai Healthcare City, P.O. Box 505108
Dubai, United Arab Emirates
Tel + 971 436 24 762
Fax + 971 436 24 761
contactuae@draeger.com

REGION ASIA / PACIFIC

Dräger Medical
South East Asia Pte Ltd
25 International Business Park
#04-27/29 German Centre
Singapore 609916, Singapore
Tel +65 6572 4388
Fax +65 6572 4399
asia.pacific@draeger.com

REGION NORTH AMERICA

Draeger Medical, Inc.
3135 Quarry Road
Telford, PA 18969-1042, USA
Tel +1 215 721 5400
Toll-free +1 800 437 2437
Fax +1 215 723 5935
info.usa@draeger.com

Manufacturer:

Dräger Medical GmbH
23542 Lübeck, Germany
The quality management system at Dräger Medical GmbH is certified according to ISO 13485, ISO 9001 and Annex II.3 of Directive 93/42/EEC (Medical devices).